

KIMBERLY-CLARK* PURPLE NITRILE-XTRA* Nitrile Exam Gloves

Product Information

Code	Description	Size
55090	KIMBERLY-CLARK* PURPLE NITRILE-XTRA* Exam Gloves 12" Ambi	X-Small
50601	KIMBERLY-CLARK* PURPLE NITRILE-XTRA* Exam Gloves 12" Ambi	Small
50602	KIMBERLY-CLARK* PURPLE NITRILE-XTRA* Exam Gloves 12" Ambi	Medium
50603	KIMBERLY-CLARK* PURPLE NITRILE-XTRA* Exam Gloves 12" Ambi	Large
50604	KIMBERLY-CLARK* PURPLE NITRILE-XTRA* Exam Gloves 12" Ambi	X-Large

Material: Acrylonitrile Butadiene, latex-free, powder-free. Tested for use with ten chemotherapy drugs

Design: Ambidextrous, beaded cuff, with textured fingertips

Packaging: 50 gloves per dispenser box. 10 Dispenser boxes per case. Shipping case and inner packaging labeled with product name, size, lot number and catalog number.

This glove meets the emergency medical examination glove requirements of the NFPA 1999, Standard On Protective Clothing for Medical Emergency Operations, 2003 Edition.



Physical Properties

Characteristics	Value	Test Method
Freedom from holes	1.5 AQL ¹	ASTM D 5151

¹ AQL as defined per ISO 2859-1 for sampling by attributes

Tensile Properties	Tensile Strength		Ultimate Elongation	ASTM D 412 and D 573
	Before Aging	21 MPa, nominal	550%, nominal	
After Accelerated Aging	21 MPa, nominal	550%, nominal		

Dimensional	Measured Point	mm		mil		ASTM D 3767 and D 6319
		0.15	5.9	0.12	4.7	
Nominal Thickness	Middle finger	0.15	5.9	0.12	4.7	
	Palm	0.12	4.7	0.09	3.5	
	Cuff	0.09	3.5			
Nominal Length	305mm / 12 inches					

Palm Widths	X-Small					Small					Medium					Large					X-Large					ASTM D 3767 and D 6319
	70	80	95	110	120	70	80	95	110	120	70	80	95	110	120	70	80	95	110	120						
Nominal Width (mm)	70	80	95	110	120	70	80	95	110	120	70	80	95	110	120	70	80	95	110	120	70	80	95	110	120	